

Invacare® Matrx®
E2 BACK (E2S/E2D)
with **FIXED HARDWARE**



INSTALLATION/ SET-UP MANUAL

i IMPORTANT! The installation instructions contained in this manual must only be carried out by a qualified service technician.

matrx®
SEATING SERIES

INVACARE
Yes, you can.®

TRD0566 Rev C

WARRANTY

PLEASE NOTE: THE WARRANTY BELOW HAS BEEN DRAFTED TO COMPLY WITH FEDERAL LAW APPLICABLE TO PRODUCTS MANUFACTURED AFTER JULY 4, 1975.

This warranty is extended only to the original purchaser/user of our products. This warranty gives you specific legal rights and you may also have other legal rights which vary from state to state. Invacare/Motion Concepts warrants this product to be free from defects in materials and workmanship for two years of use by original purchaser. This warranty does not apply to punctures, tears or burns, nor to the removable cushion cover. If within such warranty period any such product shall be proven to be defective, such product shall be repaired or replaced, at Invacare's/Motion Concepts' option, with refurbished or new parts. This warranty does not include any labor or shipping charges incurred in replacement part installation or repair of any such product. Product repairs shall not extend this warranty-coverage for repaired product shall end when this limited warranty terminates. Invacare's/Motion Concepts' sole obligation and your exclusive remedy under this warranty shall be limited to such repair and/or replacement. For warranty service, please contact the dealer from whom you purchased your Invacare/Motion Concepts product. In the event you do not receive satisfactory warranty service, please write directly to Invacare/Motion Concepts at the address on the back cover. Provide dealer's name, address, model number, the date of purchase, indicate nature of the defect and, if the product is serialized, indicate the serial number.

Invacare Corporation/Motion Concepts will issue a return authorization. The defective unit or parts must be returned for warranty inspection using the serial number, when applicable, as identification within thirty (30) days of return authorization date.

DO NOT return products to our factory without our prior consent. C.O.D. shipments will be refused; please prepay shipping charges.

LIMITATIONS AND EXCLUSIONS: THE WARRANTY SHALL NOT APPLY TO PROBLEMS ARISING FROM NORMAL WEAR OR FAILURE TO ADHERE TO THE ENCLOSED INSTRUCTIONS. IN ADDITION, THE FOREGOING WARRANTY SHALL NOT APPLY TO SERIAL NUMBERED PRODUCTS IF THE SERIAL NUMBER HAS BEEN REMOVED OR DEFACED; PRODUCTS SUBJECTED TO NEGLIGENCE, ACCIDENT, IMPROPER OPERATION, MAINTENANCE OR STORAGE; OR PRODUCTS MODIFIED WITHOUT INVACARE'S/MOTION CONCEPTS' EXPRESS WRITTEN CONSENT INCLUDING, BUT NOT LIMITED TO: MODIFICATION THROUGH THE USE OF UNAUTHORIZED PARTS OR ATTACHMENTS; PRODUCTS DAMAGED BY REASON OF REPAIRS MADE TO ANY COMPONENT WITHOUT THE SPECIFIC CONSENT OF INVACARE/MOTION CONCEPTS; PRODUCTS DAMAGED BY CIRCUMSTANCES BEYOND INVACARE'S/MOTION CONCEPTS' CONTROL; PRODUCTS REPAIRED BY ANYONE OTHER THAN AN INVACARE/ MOTION CONCEPTS DEALER, SUCH EVALUATION SHALL BE SOLELY DETERMINED BY INVACARE/ MOTION CONCEPTS.

THE FOREGOING WARRANTY IS EXCLUSIVE AND IN LIEU OF ALL OTHER EXPRESS WARRANTIES, IF ANY, INCLUDING THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. IT SHALL NOT EXTEND BEYOND THE DURATION OF THE EXPRESSED WARRANTY PROVIDED HEREIN AND THE REMEDY FOR VIOLATIONS OF ANY IMPLIED WARRANTY SHALL BE LIMITED TO REPAIR OR REPLACEMENT OF THE DEFECTIVE PRODUCT PURSUANT TO THE TERMS CONTAINED HEREIN. INVACARE/MOTION CONCEPTS SHALL NOT BE LIABLE FOR ANY CONSEQUENTIAL OR INCIDENTAL DAMAGES WHATSOEVER. THIS WARRANTY SHALL BE EXTENDED TO COMPLY WITH STATE/PROVINCIAL LAWS AND REQUIREMENTS.

This manual may not be reproduced or reprinted either partly or completely without previous written consent from Motion Concepts or its statutory representatives.


This user manual is compiled from the latest specifications and product information available at the time of publication. We reserve the right to make any necessary as they become necessary. Any changes made to our products may cause slight variations between the illustrations and explanations in this manual and the product you have purchased. If you have any questions or concerns regarding the information provided in this manual, please contact our Customer Service Department for assistance.


1.0 GENERAL WARNINGS/INFORMATION

1.1 OVERVIEW

SAFETY SYMBOLS:

Proper set-up and safe use of your Invacare Matrix E2S/E2D Back depends on your own good judgement and common sense, as well as that of your provider, caregiver and/or health professional. The following symbols are used throughout this user manual to identify warnings, cautions and important notes. It is very important that you read and understand them completely.

 **CAUTION!** Failure to heed the cautions in this manual may result in damage to your Matrix Seating product.

 **WARNING!** Failure to heed the warnings in this manual may result in death, serious injury or damage.


 **IMPORTANT!** Important information to remember during installation, set-up and/or adjustment of you Matrix Seating product.

USER WARNING!

Improper use of this product may cause injury or damage. DO NOT use this product or any available optional equipment without first completely reading and understanding these instructions and any additional instructional material such as user manual, service manuals or instruction sheets supplied with this product or optional equipment. If you are unable to understand the warnings, cautions or instructions, contact a health care professional or dealer before attempting to use this equipment.

IMPORTANT INFORMATION

The best way to avoid problems related to pressure injuries is to understand their causes and your role in a skin management program. Your therapist and physician should be consulted if you have questions regarding individual limitations and needs. All back supports should be selected carefully. Working with your therapist and physician is the best way to assure that a cushion choice matches your individual needs. As the needs of the individual become more complex, the evaluation becomes more important.

 **IMPORTANT!** Please refer to the Wheelchair Owners/Operators Manual for additional safety information, warnings and cautions related to the operation of your wheelchair.

1.2 INTENDED USE

Our Invacare® Matrix® E2 backs are intended to provide back support for a wide variety of wheelchair users. It is designed to allow precise orientation within a wheelchair to ensure optimal postural support and positioning. Our E2 Backs can be installed on a wide range of wheelchairs.

2.0 SAFETY INFORMATION

2.1 TRANSPORTATION WARNINGS!

▲ WARNING! Risk of Death, Serious Injury or Damage

- Risk of severe injury when transporting the client within a wheelchair fitted with an Invacare® Matrx® E2 Back in a motor vehicle.

TRANSPORTING THE CLIENT IN A MOTOR VEHICLE WITHIN A WHEELCHAIR FITTED WITH A MATRX BACK

The following guidelines should be made available to all parties responsible for the transportation of the client, such as schools and transport providers.

- Wherever practicable Motion Concepts recommends that the client is transferred out of the wheelchair into a vehicle seat using a lap and diagonal seat belt or an appropriate child safety seat.
- The following back supports have been dynamically tested and meet the requirements of ISO 16840-4:2009, Wheelchair Seating- Part 4: Seating Systems for Use in Motor Vehicles; and Section 20, ANSI/RESNA WC-4:2017, Wheelchair Seating Systems for Use in Motor Vehicles:
 - Invacare® Matrx® E2S back with Removable Hardware and Elan Headrest
 - Invacare® Matrx® E2S back with Fixed Hardware and Elan Headrest

A forward facing frontal impact test was performed for each configuration using an ISO/RESNA Surrogate Wheelchair Frame (SWCF) and a Hybrid III Midsize-Adult Male Anthropomorphic Test Device (ATD). The SWCF was secured by a four-point, strap type wheelchair tie-down and the ATD was restrained by a surrogate three-point belt with a SWCF-Anchored Lap Belt.

- Invacare® Matrx® E2 Backs used to transport a client should be equipped with a Headrest Support (i.e. restricted to back heights 14in (36cm) and higher); and may **only** be used in conjunction with wheelchairs that comply with the performance requirements of RESNA WC19 or ISO 7176-19.
- For safe transportation of a client within a wheelchair in a motor vehicle, refer to the user manuals of the wheelchair and tie-down and occupant restraint system (WTORS)- see additional information below.

BEFORE ENTERING THE VEHICLE CHECK THE FOLLOWING:

The Matrx Back must be securely fastened into the wheelchair as described in the user manual.

Posture Belts & Harnesses should remain fitted during transportation, although they are designed to provide postural support only, they are NOT intended for safety during transportation.

A Head Support should be securely fitted and correctly adjusted close to the rear of the client's head during transportation, to reduce the risk of whiplash injuries.

WHEELCHAIR TIE-DOWN & OCCUPANT RESTRAINT SYSTEM (WTORS)

- **Wheelchair Tie-down:** The wheelchair must be securely tied down within the vehicle, in a forward facing position, in accordance with the wheelchair and tie-down manufacturers instructions.
- **Occupant Restraint:** A suitable occupant restraint system must also be fitted in accordance with the manufacturer's instructions. A minimum is a lap and diagonal

belt (lap belts on their own are not suitable). Restraints that loop over the shoulder and anchor to the vehicle floor should be avoided where possible, as they can cause a heavy downward load through the client during an impact, the preferred type anchor above and behind the shoulder as with a car seat belt. The lap section of the belt should fit snugly over the pelvis of the client and should not be able to ride up to the abdominal area. The upper torso section of belt should be in contact with the client's chest and fit over the shoulder(s), while not cutting into the neck or slipping off the shoulder(s).

i IMPORTANT! If the wheelchair and Invacare® Matrx® E2 Back is transported unoccupied, the wheelchair should be tied down with suitable equipment according to the manufacturer's instructions. If the equipment is dismantled for transportation, ensure that all parts are secured safely within the vehicle.

2.2 GENERAL SAFETY AND INSTALLATION WARNINGS!

⚠ WARNING! Risk of Death, Serious Injury, or Damage

Lighted cigarettes dropped onto an upholstered seating system can cause a fire resulting in death, serious injury or damage. Wheelchair occupants are at particular risk of death or serious injury from these fires and resulting fumes because they may not have the ability to move away from the wheelchair.

- DO NOT smoke while using this wheelchair

⚠ WARNING! Risk of Serious Injury

When establishing the mounting position for the Invacare® Matrx® back, ensure the stability of the wheelchair is not compromised. Adjusting the forward and aft position and/or angle of the backrest will alter the user's center of gravity within the wheelchair. It may be necessary to reposition the front casters, rear wheels, back angle, tilt-in-space, recline position and/or seat depth before use.

⚠ WARNING! Risk of Serious Injury or Damage

- The procedures in this manual should be performed by a qualified technician
- The hardware provided is high strength and tested for durability. **Do Not substitute hardware.** Use only the original hardware supplied with your Matrx back
- Attaching hardware that is loosely secured could cause loss of stability resulting in serious injury or damage.
- After ANY adjustments, repair or service and before use, make sure that all attaching hardware is tightened securely.
- DO NOT install the Invacare® Matrx® back assembly onto back canes with an outside diameter greater than 1" (25mm) or less than 3/4" (19mm). Otherwise, injury or damage may occur.
- For power seating applications, the Fixed Mounting Hardware must be used.

⚠ WARNING! Risk of Injury

Invacare® Matrx® products are specifically designed and manufactured for use in conjunction with Invacare® Matrx® accessories. Accessories designed by other manufacturers have not been tested by Motion Concepts and are not recommended.

⚠ WARNING Risk of Injury

Skin condition should be checked very frequently after the installation of any new seating system. Working with you therapist, physician and service provider is the best way to assure that a seating choice meets your individual needs.


3.0 MATRX E2 BACK INSTALLATION

3.1 FITTING THE E2 BACK TO THE WHEELCHAIR:

PRIOR to installation, verify the diameter of the wheelchair back canes (A) and the width between back canes (B)


Model #	Description	Contour Depth
E2S wwhh	Matrx E2 Back	3in (8cm)
E2D wwhh	Matrx E2 Deep Back	6in (15cm)

*wwhh = width/height



*requires cane inserts (provided)

E2S/E2D Back Width	FITS ONTO BACK CANE WIDTH (B)		
	Min Width	MiniSet Max Width	EasySet Max Width
14in (36cm)	13in (33cm)	15in (38cm)	16in (41cm)
15in (38cm)	14in (36cm)	16in (41cm)	17in (43cm)
16in (41cm)	15in (38cm)	17in (43cm)	18in (46cm)
17in (43cm)	16in (41cm)	18in (46cm)	19in (48cm)
18in (46cm)	17in (43cm)	19in (48cm)	20in (51cm)
19in (48cm)	18in (46cm)	20in (51cm)	21in (53cm)
20in (51cm)	19in (48cm)	21in (53cm)	22in (56cm)



The type of Mounting Hardware can vary depending on several factors, including the height of the back shell, the needs of the end user and the type of wheelchair.

E2S/E2D Back Height	STANDARD Hardware	OPTIONAL Mounting Hardware			Weight Capacity
	Removable	Heavy-Duty	Fixed	Split	
6in (15cm)	MiniSet -see Installation Instruction TRD0552	HD MiniSet -see Installation Instruction TRD0552	FIXED MiniSet -see Installation Instruction TRD0566	N/A	300lb (136kg)
8in (20cm)					
10in (26cm)					
12in (31cm)*					
14in (36cm)*					
16in (41cm)*					
18in (46cm)	EasySet -see Installation Instruction TRD0552	HD EasySet -see Installation Instruction TRD0552	FIXED EasySet -see Installation Instruction TRD0566	SPLIT Hardware -see Installation Instruction TRD0247	300lb (136kg)
20in (51cm)					

***Note:** E2S/E2D Backs may be configured with EasySet hardware for the back heights indicated. (The range of back height adjustment will be limited for these configurations)

3.2 E2 FIXED HARDWARE ASSEMBLY OVERVIEW

i IMPORTANT! The E2 Fixed hardware utilizes a unique **fixed mounting plate*** in place of the EasySet/MiniSet mounting plate and mounting bracket assembly used on the E2 Removable Hardware. If converting from removable to fixed hardware, the cane clamps from the E2 Removable Hardware assembly are reused to complete the E2 Fixed Hardware installation.

***Note:** There are two styles of fixed mounting plates for use on the E2 Back (depending on the overall height of the back shell). Back heights of 16" (41cm) or less come standard with **Fixed MiniSet** hardware; and back heights of 18" (46cm) and greater come standard with **Fixed EasySet** hardware

3.2.1 FIXED MINISet HARDWARE

Standard Mounting Hardware Kit for Back Heights **16" (41cm) or less**

2 - MiniSet Fixed Mounting Plate (A)

4 - Width/Height Adjustment Screws (B)

2 - E2 MiniSet Washer Plate (C)

2 - E2 MiniSet Nut Plate (D)

2 - Cane Clamp Ass'y (E)


2 - Cane Inserts 7/8" (22mm) (F)

2 - Cane Inserts 3/4" (19mm) (G)


4 - Adjustment Screws/Washers (H)

2 - Cane Clamp Screw/Washer (I)

2 - Cane Clamp Nut Plate (J)



cane inserts (if applic.)



**Fixed MiniSet Hardware
Assembly (Installed)**

3.2.2. FIXED EASYSET HARDWARE

Standard Mounting Hardware Kit for Back Heights **18" (46cm) or greater**
(¹Optional Mounting Hardware Kit for Back Heights 12/14/16" (31/36/41cm))

2 - EasySet Fixed Mounting Plate (A)

4 - Width/Height Adjustment Screws (B)

2 - E2 EasySet Washer Plate (C)

2 - E2 EasySet Nut Plate (D)

2 - Cane Clamp Assy'y (E)

2 - Cane Inserts 7/8" (22mm) (F)



2 - Cane Inserts 3/4" (19mm) (G)

6 - Adjustment Screws*/Washers (H)


6 - Cane Clamp Screws/Washers (I)

2 - Cane Clamp Nut Plate (J)

^{*}The Fixed EasySet hardware kit uses **shorter** length Depth/Angle Adjustment Screws (H)



¹**Note:** The range of back height adjustment will be limited/reduced when using these non-standard (optional) hardware configurations.



i **NOTE:** If any of the supplied hardware is missing or misplaced, please contact your local Service Provider to obtain the necessary replacements.

3.3 E2 FIXED MOUNTING PLATE INSTALLATION

i IMPORTANT! Unless otherwise necessary, **always** ensure that adjustments are made **symmetrically** on the left and right mounting brackets.

i IMPORTANT! The **Fixed MiniSet** mounting plates allow for installation of the E2S/E2D back onto back canes up to **1" (25mm) wider or 1" (25mm) narrower** than the nominal back width. The **Fixed EasySet** mounting plates allows for installation of the E2 Back onto back canes up to **2" (51mm) wider or 1" (25mm) narrower** than the nominal back width.

3.3.1 INSTALLING THE E2 FIXED MOUNTING PLATES:

i NOTE: Prior to installing the Fixed Mounting Brackets, the E2 Back Cover must first be removed by releasing the hook and loop fasteners along the bottom edge and inside of the back shell.

Fixed Mounting Plate Installation: (see Fig. 1 and Fig. 2)

1. Working on one side of the back shell at a time, align the washer plate (B) with the slots in the fixed mounting plate (A) and insert the adjustment screws (C) through both parts.
2. From the rear of the back shell, align the mounting plate assembly (from step 1 above) with the vertical slot in the E2 back shell, then secure* the adjustment screws into the threaded nut plate (on the inside of the back shell).
3. Repeat steps 1. and 2. to install the fixed mounting bracket onto the opposite side of the back shell.

(*note: the fixed mounting plates can remain loosely installed at this time. The mounting plate assemblies will be fully secured following final installation/adjustment of the E2 Back- refer to **section 3.6**)

Fig. 1. MiniSet Fixed Mounting Plate Installation





Fig. 2. EasySet Fixed Mounting Plate Installation






3.4 CANE CLAMP ORIENTATION

The orientation of the Cane Clamps (A) provides approximately **4" (102mm)** of fore/aft adjustment for the **Fixed EasySet** hardware, and **3" (76mm)** of fore/aft adjustment for the **Fixed MiniSet** hardware. PRIOR to installing the Cane Clamps, determine the orientation that will best meet the needs of the end user.

- **Standard Orientation (Fig. 3a/4a) = maximized seat depth**
- **Inverted Orientation (Fig. 3b/4b) = minimized seat depth**

NOTE: Additional depth adjustments can be made via the slotted channels in the fixed mounting plates - see **section 3.6.2**.







(*Requires the mounting plate to be disassembled from the cane clamp and rotated 180°)


3.5 CANE CLAMP INSTALLATION (see Fig. 5a/5b; Fig 6a/6b)

i NOTE: The mounting position of the cane clamps should be estimated so that, when installed, the installed E2 back will be close to the desired back height. A higher cane clamp mounting position typically works best. Additional fine-tune height adjustment of the back shell will be available following installation

1. (If applicable) Remove existing wheelchair back according to the wheelchair manufacturers instructions.
2. Install the MiniSet **or** EasySet Cane Clamps onto each back cane as follows:
 - (If necessary) Install the appropriate cane insert onto the back cane (1).
 - Position/close the outer clamp (A) over the back cane (cane insert) and align it with the inner clamp (B).
 - Loosely secure the clamp assembly to the back cane via the mounting screw(s) (2).
 - Adjust the height* of each clamp assembly (3) equally along the back canes in order to achieve the estimated final back height.
 - Fully secure all cane clamp mounting screws (2)

***NOTE:** The final back height can be further adjusted (via mounting brackets) following installation of the E2 Back- see **section 3.6.1**





3.6 E2 BACK INSTALLATION, SET-UP & ADJUSTMENT

NOTE: During final set-up, we recommend having the end user seated in the wheelchair to best fit/position the E2S/E2D back.

IMPORTANT! Unless otherwise necessary, **always** ensure that adjustments are made **symmetrically** on the left and right mounting brackets.

PLEASE NOTE

The E2 Fixed Mounting Plate screws (A) will simultaneously adjust the **Width and Height** of the back (relative to the back canes). Please read through ALL set-up/adjustment instructions, **PRIOR** to beginning the back installation.

The vertical (height) adjustment slots on each E2 back shell will vary in length depending on the nominal back height- refer to **Table 1**.

Table 1. BACK SHELL: RANGE OF HEIGHT ADJUSTMENT

Back Height	Adjustment Range	Back Height	Adjustment Range
6" (15cm)	1.25" (32mm)	12" (31cm)	2.5" (63mm)
8" (23cm)	1.5" (38mm)	14", 16" (36, 41cm)	3.5" (89mm)
10" (26cm)	2.0" (51mm)	18", 20" (46, 51cm)	3.25" (83mm)

3.6.1 BACK INSTALLATION, WIDTH AND HEIGHT ADJUSTMENT:

Installation, Width and Height Adjustment: (see Fig. 7a/7b, 8a/8b)

1. Loosen (do not remove) the adjustment screws (A), then adjust the width of both mounting plates (B) horizontally (1) so that the mounting plates align with the left and right cane clamp assemblies- see Fig 7a/8a.
2. To install the E2 Back (2), insert the mounting screws (C) through the cane clamp (E) and mounting plate (B), and secure into the threaded nut plate (D).
3. With the E2 Back securely installed, set the final back height* and adjust the back shell so that it is centered (symmetrically) between the back canes (3). When possible, this step should be performed with the user seated in the chair in order to best set the back position.

NOTE: use the hash marks along the vertical slots on the back shell to ensure both back height (mounting brackets) are adjusted equally. The minimum recommended back height* is 4"- 5" (10-13cm) from the top of the seat surface - see Fig 7b/8b.

4. Tighten the adjustment screws (A) on both mounting plates (B) to secure the E2 Back in the desire position.
5. Refer to Section 3.6.2 for E2 Back Depth and Angle Adjustments.

***Note:** The minimum recommended back height is provided as a guideline only. The final back height should be adjusted/set to best meet the end user's needs-see **Important!** note below.

i IMPORTANT! Always ensure that the bottom of the back is positioned anatomically below the Posterior Superior Iliac Spine (PSIS). If you have any questions or concerns regarding proper back height adjustment, please contact your Motion Concepts representative for assistance.

Fig. 7a E2 Back Installation- Fixed MiniSet:




Fig. 7b Back Height/Centering Adjustment- Fixed MiniSet:




Fig. 8a Back Installation- Fixed EasySet:





Fig. 8b Back Height/Centering Adjustment- Fixed EasySet:



3.6.2 BACK DEPTH AND ANGLE ADJUSTMENT:


! PLEASE NOTE

The Mounting Screws (C) on the cane clamp assembly will simultaneously adjust the **Depth and Angle** of the E2 Back (relative to the back canes). Please read through ALL adjustment instructions, PRIOR to beginning the adjustments


1. With the end user seated in the chair (when possible), loosen (do not remove) the mounting screws (C) on the left and right cane clamp assemblies, then adjust the E2 back depth (Fig. 9a/10a) (4) and back angle (Fig. 9b/10b) (5), as necessary, to best fit the needs of the end user*.
2. Once the final back depth and angle has been determined, fully tighten the mounting screws (C) on both cane clamps (E) to secure the E2 Back into place, and verify ALL mounting/adjustment hardware is fully secured.

**note: the slotted channels in the fixed mounting plates allow for approximately 1" (26mm) of depth adjustment on the Fixed MiniSet hardware, and up to 1.5" (39mm) of depth adjustment on the Fixed EasySet hardware. Both fixed hardware assemblies allow for approximately $\pm 20^\circ$ of angle adjustment.*

Back Depth and Angle Adjustment - Fixed MiniSet



Back Depth and Angle Adjustment - Fixed EasySet



i **NOTE:** The range of depth adjustment will be reduced when the angle of the E2 Back (fixed mounting plate) is altered.

4.0 MAINTENANCE

4.1 INSPECTION AND MAINTENANCE

WARNING! Risk of injury or damage

- Do not continue to use this product if problems or damage are discovered;
- Corrective maintenance can be performed at or arranged through your service provider

Weekly: Inspect **ALL** fasteners weekly to ensure that mechanical connections and attaching hardware are tightened securely.

Monthly: Perform a visual inspection of parts including hardware, brackets, upholstery materials, foam, and plastics for deformation, corrosion, breakage, wear or compression. Replace damaged/worn components if necessary.

4.3 REUSE/RECYCLING

WARNING! Risk of Injury

Always have an authorized dealer inspect the product for damage or wear PRIOR to transferring the product to a different user. If any damage is found, DO NOT use this product. Failure to heed these warnings may result in severe injury.

This product is suitable for re-use. The outer cover (fabric and foam), back shell and mounting hardware must be inspected by a qualified dealer or healthcare professional before re-use. If any damage is found, DO NOT RE-USE

Clean and disinfect all components (cover, foam, hardware and back shell) in accordance with local government regulations, prior to re-use. For further assistance and/or information on decontamination and disinfection, please consult your local Infection Control Office.

4.4 DISPOSAL

The disposal and recycling of used devices and packaging must comply with the applicable local legal regulations.

If you have any concerns or questions regarding the set-up and adjustment of the Matrix E2 Back, please contact our Technical Service Department for assistance.


USA: 888-433-6818

CAN: 800-680-4191

5.0 E2 BACK ACCESSORIES

5.1 E2 HEADREST INTERFACE PLATE

The E2 Headrest Interface Plate (A) is included only with Matrix E2 Backs for **Back Heights** ranging from **14" (36cm) up to 20" (51cm)**. The E2 headrest interface is compatible with most industry standard headrest mounting hardware, including our **Matrix Elan Headrest Hardware** (sold separately). The headrest interface plate installs directly onto the E2S/E2D back shell using the hardware provided*.



***NOTE:** Headrest mounting hardware must be installed onto the E2 Interface Plate **PRIOR** to installation onto the back shell

5.2 E2 LUMBAR SUPPORT

The E2 Lumbar Support (A) is included with ALL E2 Backs. The Lumbar Support is a pre-fabricated foam insert that provides optional lower back support for improved positioning and comfort. The lumbar support may be installed inside the Matrix E2S/E2D cover when desired (see instructions/images below).


Installing the Lumbar Support:

1. Open the outer cover at the bottom rear of the back shell (via the hook and loop fastener).
2. Install the lumbar support inside the E2 back cover (behind the existing foam pad), as illustrated below.
3. Refasten the outer cover onto the back shell.



5.3 ADDITIONAL E2 ACCESSORY OPTIONS (Sold Separately)

- I. PRIVACY FLAP (PFLAP)
- II. BACKBONE (PSIS/SacroLumbar Support)
- III. POSTURE FLEX (Abdominal Support) and QUAD RELEASE (QR) TAB
- IV. CHEST HARNESS INTERFACE
- V. EASYFIX LATERALS
- VI. SWING-AWAY LATERALS





Invacare

www.invacare.com**USA**

One Invacare Way
Elyria, Ohio USA
44036-2125
440-329-6000
800-333-6900

Canada

570 Matheson Blvd. E Unit B
Mississauga, Ontario
L4Z 4G4 Canada
800-668-5324



Motion Concepts

www.motionconcepts.com**USA**

700 Ensminger Rd. Suite 112
Tonawanda, NY 14150
888-433-6818

Canada

84 Citation Dr., Unit 1-10
Concord, Ontario L4K 3C1
866-748-7943